

# Piccolo<sup>®</sup> Electrolyte Panel



For In Vitro Diagnostic Use and For Professional Use Only

Customer and Technical Service: 800-822-2947

Customers outside the US should contact their local  
Abaxis representative for customer service

Applicable to US customers only

**CLIA Waived:** Use lithium heparin whole blood, only  
**Moderate Complexity:** Use lithium heparin whole blood,  
lithium heparin plasma, or serum

April 2014

PN: 400-7116-1 Rev. P

## 1. Intended Use

The Piccolo<sup>®</sup> Electrolyte Panel, used with the Piccolo<sup>®</sup> blood chemistry analyzer or the Piccolo Xpress<sup>®</sup> chemistry analyzer, is intended to be used for the *in vitro* quantitative determination of chloride, potassium, sodium, and total carbon dioxide in heparinized whole blood, heparinized plasma, or serum in a clinical laboratory setting or point-of-care location.

### For US Customers Only

The tests on this panel are waived under CLIA '88 regulations. If a laboratory modifies the test system instructions, then the tests are considered high complexity and subject to all CLIA requirements. For CLIA waived labs, only lithium heparin whole blood may be tested. For use in moderate complexity labs, lithium heparinized whole blood, lithium heparinized plasma, or serum may be used.

A CLIA Certificate of Waiver is needed to perform CLIA waived testing. A Certificate of Waiver can be obtained from the Centers for Medicare & Medicaid Services (CMS). Please contact the Commission on Laboratory Accreditation (COLA) at 1-800-981-9883 for assistance in obtaining one.

## 2. Summary and Explanation of Tests

The Piccolo Electrolyte Panel and the Piccolo blood chemistry analyzer comprise an *in vitro* diagnostic system that aids the physician in diagnosing the following disorders:

|                       |  |
|-----------------------|--|
| Chloride:             | Dehydration, prolonged diarrhea and vomiting, renal tubular disease, hyperparathyroidism, burns, salt-losing renal diseases, overhydration and thiazide therapy.   |
| Potassium:            | Renal glomerular or tubular disease, adrenocortical insufficiency, diabetic ketacidosis, excessive intravenous potassium therapy, sepsis, panhypopituitarism, in vitro hemolysis, hyperaldosteronism, malnutrition, hyperinsulinism, metabolic alkalosis and gastrointestinal loss.                  |
| Sodium:               | Dehydration, diabetes insipidus, loss of hypotonic gastrointestinal fluids, salt poisoning, selective depression of sense of thirst, skin losses, burns, sweating, hyperaldosteronism, CNS disorders, dilutional, depletion and delusional hyponatremia and syndrome of inappropriate ADH secretion. |
| Total carbon dioxide: | Primary metabolic alkalosis and acidosis and primary respiratory alkalosis and acidosis.   |


**As with any diagnostic test procedure, all other test procedures including the clinical status of the patient, should be considered prior to final diagnosis.**

## 3. Principle of Procedure

### **Chloride (CL<sup>-</sup>)**

The method is based on the determination of chloride-dependent activation of  $\alpha$ -amylase activity. Deactivated  $\alpha$ -amylase is reactivated by addition of the chloride ion, allowing the calcium to re-associate with the enzyme. The reactivation of  $\alpha$ -amylase activity is proportional to the concentration of chloride ions in the sample. The reactivated  $\alpha$ -amylase converts the substrate, 2-chloro-*p*-nitrophenyl- $\alpha$ -D-maltotriose (CNP3) to 2-chloro-*p*-nitrophenol (CNP) producing color and


$\alpha$ -maltotriose (G3). The reaction is measured bichromatically and the increase in absorbance is directly proportional to the reactivated  $\alpha$ -amylase activity and the concentration of chloride ion in the sample.<sup>1</sup>



### Potassium ( $\text{K}^+$ )

Spectrophotometric methods have been developed that allow the measurement of potassium concentration on standard clinical chemistry instrumentation. The Abaxis enzymatic method is based on the activation of pyruvate kinase with potassium and shows excellent linearity and negligible susceptibility to endogenous substances.<sup>2,3,4</sup> Interference from sodium and ammonium ion are minimized with the addition of Kryptofix and glutamine synthetase, respectively.<sup>2</sup>


In the coupled-enzyme reaction, pyruvate kinase (PK) dephosphorylates phosphoenolpyruvate (PEP) to form pyruvate. Lactate dehydrogenase (LDH) catalyzes conversion of pyruvate to lactate. Concomitantly, NADH is oxidized to  $\text{NAD}^+$ .



The rate of change in absorbance difference between 340 nm and 405 nm is due to the conversion of NADH to  $\text{NAD}^+$  and is directly proportional to the amount of potassium in the sample.

### Sodium ( $\text{Na}^+$ )


Colorimetric and enzymatic methods have been developed that allow the measurement of sodium concentration on standard clinical chemistry instrumentation.<sup>5,6,7</sup> In the Abaxis enzymatic reaction,  $\beta$ -galactosidase is activated by the sodium in the sample. The activated enzyme catalyzes the reaction of o-nitrophenyl- $\beta$ -D-galactopyranoside (ONPG) to o-nitrophenol and galactose.



### Total Carbon Dioxide ( $\text{tCO}_2$ )

Total carbon dioxide in serum or plasma exists as dissolved carbon dioxide, carbamino derivatives of proteins, bicarbonate and carbonate ions and carbonic acid. Total carbon dioxide can be measured by pH indicator,  $\text{CO}_2$  electrode and spectrophotometric enzymatic methods, which all produce accurate and precise results.<sup>8,9</sup> The enzymatic method is well suited for use on a routine blood chemistry analyzer without adding complexity.

In the enzymatic method, the specimen is first made alkaline to convert all forms of carbon dioxide ( $\text{CO}_2$ ) toward bicarbonate ( $\text{HCO}_3^-$ ). Phosphoenolpyruvate (PEP) and  $\text{HCO}_3^-$  then react to form oxaloacetate and phosphate in the presence of phosphoenolpyruvate carboxylase (PEPC). Malate dehydrogenase (MDH) catalyzes the reaction of oxaloacetate and reduced nicotinamide adenine dinucleotide (NADH) to  $\text{NAD}^+$  and malate. The rate of change in absorbance due to the conversion of NADH to  $\text{NAD}^+$  is directly proportional to the amount of  $\text{tCO}_2$  in the sample.



## 4. Principle of Operation

See the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual, for the Principles and Limitations of the Procedure.

## 5. Description of Reagents

### Reagents

Each Piccolo Electrolyte Panel contains dry test-specific reagent beads (described below). A dry sample blank reagent (comprised of buffer, surfactants, excipients, and preservatives) is included in each disc for use in calculating concentrations of chloride ( $\text{CL}^-$ ), potassium ( $\text{K}^+$ ), sodium ( $\text{NA}^+$ ), and total carbon dioxide ( $\text{tCO}_2$ ). Each disc also contains a diluent consisting of surfactants and preservatives.

**Table 1: Reagents**

| Component  | Quantity/Disc |
|--|---------------|
| Adenosine-5'-diphosphate   | 13.9µg        |
| Amylase  | 0.036U        |
| Calcium acetate  | 25.2µg        |
| 2-Chloro-4-nitrophenyl – alpha-maltotrioxide (CNPG3)                       | 52.5µg        |
| Citric acid, trisodium salt  | 567µg         |
| Ethylenediaminetetraacetic acid (EDTA)                                     | 182µg         |
| Ethylene glycol-bis(β-aminoethyl ether)-N,N,N',N'-tetraacetic acid (EGTA)  | 3.7µg         |
| β-Galactosidase  | 0.0046U       |
| 4,7,13,16,21,24-Hexaoxa-1,10-diazabicyclo[8.8.8]hexacosane (Kryptofix 222) | 0.3µg         |
| Lactate dehydrogenase  | 0.3U          |
| Magnesium acetate  | 15µg          |
| Malate dehydrogenase (porcine heart)                                       | 0.1U          |
| N-acetyl cysteine  | 15.3µg        |
| β-Nicotinamide adenine dinucleotide, reduced (NADH)                        | 21µg          |
| α-Oxoglutarate   | 7.9µg         |
| 4,7,13,16,21-Pentaoxa-1,10-diazabicyclo[8.8.5]tricosane (Kryptofix 221)    | 84µg          |
| Phosphoenol pyruvate   | 34µg          |
| Phosphoenol pyruvate carboxylase   | 0.001U        |
| Pyruvate kinase  | 0.01U         |
| Buffers, surfactants, excipients and preservatives                         |               |

### Warnings and Precautions

- For *In vitro* Diagnostic Use
- The diluent container in the reagent disc is automatically opened when the analyzer drawer closes. A disc with an opened diluent container can not be re-used. Ensure that the sample or control has been placed into the disc before closing the drawer.
- Used reagent discs contain human body fluids. Follow good laboratory safety practices when handling and disposing of used discs.<sup>10</sup> See the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual for instructions on cleaning biohazardous spills.
- The reagent discs are plastic and may crack or chip if dropped. **Never** use a dropped disc as it may spray biohazardous material throughout the interior of the analyzer.
- Reagent beads may contain acids or caustic substances. The operator does not come into contact with the reagent beads when following the recommended procedures. In the event that the beads are handled (e.g., cleaning up after dropping and cracking a reagent disc), avoid ingestion, skin contact, or inhalation of the reagent beads.

### **Instructions for Reagent Handling**

Reagent discs may be used directly from the refrigerator without warming. Do not allow discs sealed in their foil pouches to remain at room temperature longer than 48 hours prior to use. Open the sealed foil pouch, remove the disc and use according to the instructions provided in the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual. A disc not used within 20 minutes of opening the pouch should be discarded.

### **Storage**

Store reagent discs in their sealed pouches at 2-8°C (36-46°F). Do not expose opened or unopened discs to direct sunlight or temperatures above 32°C (90°F). Reagent discs may be used until the expiration date included on the package. The expiration date is also encoded in the bar code printed on the bar code ring. An error message will appear on the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Display if the reagents have expired.

### **Indications of Reagent Disc Instability/Deterioration**

A torn or otherwise damaged pouch may allow moisture to reach the unused disc and adversely affect reagent performance. Do not use a disc from a damaged pouch.

## **6. Instrument**

See the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual for complete information on use of the analyzer.

## **7. Sample Collection and Preparation**

Sample collection techniques are described in the "Sample Collection" section of the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual.

- The minimum required sample size is ~100 µL of heparinized whole blood, heparinized plasma, serum or control material. The reagent disc sample chamber can contain up to 120 µL of sample.
- Whole blood samples obtained by venipuncture must be homogeneous before transferring a sample to the reagent disc. Gently invert the collection tube several times just prior to sample transfer. Do not shake the collection tube; shaking may cause hemolysis.
- Hemolysis may cause erroneously high results in potassium assays. This problem may go undetected when analyzing whole blood (release of potassium from as few as 0.5% of the erythrocytes can increase the potassium serum level by 0.5 mmol/L). In addition, even unhemolyzed specimens that are not promptly processed may have increased potassium levels due to intracellular potassium leakage.<sup>11</sup>
- Whole blood venipuncture samples should be run within 60 minutes of collection.<sup>12</sup> The sample may be separated into plasma or serum and stored in capped sample tubes at 2-8°C (36-46°F), if the sample cannot be run within 60 minutes.
- Use only lithium heparin (green stopper) evacuated specimen collection tubes for whole blood or plasma samples. Use no-additive (red stopper) evacuated specimen collection tubes or serum separator tubes (red or red/black stopper) for serum samples.
- Start the test within 10 minutes of transferring the sample into the reagent disc.
- The concentration of total carbon dioxide is most accurately determined when the assay is done immediately after opening the tube and as promptly as possible after collection and processing of the blood in the unopened tube. Ambient air contains far less carbon dioxide than does plasma, and gaseous dissolved carbon dioxide will escape from the specimen into the air, with a consequent decrease in carbon dioxide value of up to 6 mmol/L in the course of 1 hour.<sup>13</sup>

## **8. Procedure**

### **Materials Provided**

- One Piccolo Electrolyte Panel PN: 400-1022 (a box of discs PN: 400-0022)

## Materials Required but not Provided

- Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer
- Sample transfer pipettes (fixed volume approximately 100 µL) and tips are provided with each Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer and may be reordered from Abaxis.
- Commercially available control reagents recommended by Abaxis (contact Abaxis Technical Support for approved control materials and expected values).
- Timer

## Test Parameters

The Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer operates at ambient temperatures between 15°C and 32°C (59-90°F). The analysis time for each Piccolo Electrolyte Panel is less than 14 minutes. The analyzers maintain the reagent disc at a temperature of 37°C (98.6°F) over the measurement interval.

## Test Procedure

The complete sample collection and step-by-step operating procedures are detailed in the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual.

## Calibration

The Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer is calibrated by the manufacturer before shipment. The bar code printed on the bar code ring provides the analyzer with disc-specific calibration data. See the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual.

## Quality Control

See Section 2.4 of the Piccolo blood chemistry analyzer Operator's Manual or Section 6 (Calibration and Quality Control) of the Piccolo Xpress chemistry analyzer Operator's Manual. Performance of the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer can be verified by running controls. For a list of approved quality control materials with acceptance ranges, please contact Abaxis Technical Support. Other human serum or plasma-based controls may not be compatible. Quality control materials should be stored as per the package-insert included with the controls.

If control results are out of range, repeat one time. If still out of range, call Technical Support. Do not report results if controls are outside their labeled limits. See the Piccolo blood chemistry analyzer or Piccolo Xpress chemistry analyzer Operator's Manual for a detailed discussion on running, recording, interpreting, and plotting control results.

**Waived Laboratories:** Abaxis recommends control testing as follows:

- at least every 30 days
- whenever the laboratory conditions have changed significantly, e.g. Piccolo moved to a new location or changes in temperature control
- when training or retraining of personnel is indicated
- with each new lot (CLIA waived tests in waived status labs)

**Non-Waived Laboratories:** Abaxis recommends control testing to follow federal, state, and local guidelines.

## 9. Results

The Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer automatically calculates and prints the analyte concentrations in the sample. Details of the endpoint and rate reaction calculations are found in the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual.

Interpretation of results is detailed in the Operator's Manual. Results are printed onto result cards supplied by Abaxis. The result cards or tapes have an adhesive backing for easy placement in the patient's files.

## 10. Limitations of Procedure

General procedural limitations are discussed in the Piccolo blood chemistry analyzer Operator's Manual or the Piccolo Xpress chemistry analyzer Operator's Manual.

- The only anticoagulant **recommended for use** with the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer is **lithium heparin**. Abaxis has performed studies demonstrating that EDTA, fluoride, oxalate, and any anticoagulant containing ammonium ions will interfere with at least one chemistry contained in the Piccolo Electrolyte Panel. Do not use sodium heparin.
- Samples with hematocrits in excess of 62-65% packed red cell volume (a volume fraction of 0.62-0.65) may give inaccurate results. Samples with high hematocrits may be reported as hemolyzed. These samples may be spun down to get plasma then re-run in a new reagent disc.
- **Any result for a particular test that exceeds the assay range should be analyzed by another approved test method or sent to a referral laboratory. Do not dilute the sample and run it again on the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer.**

**Warning:** Extensive testing of the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer has shown that, in very rare instances, sample dispensed into the reagent disc may not flow smoothly into the sample chamber. Due to the uneven flow, an inadequate quantity of sample may be analyzed and several results may fall outside the reference ranges. The sample may be re-run using a new reagent disc.

### Interference

Substances were tested as interferents with the analytes. Human serum pools were prepared. The concentration at which each potential interferent was tested was based on the testing levels in CLSI EP7-P.<sup>14</sup>

### Effects of Endogenous Substances

- Physiological interferents (hemolysis, icterus and lipemia) cause changes in the reported concentrations of some analytes. The sample indices are printed on the bottom of each result card to inform the operator about the levels of interferents present in each sample.
- The Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer suppresses any results that are affected by >10% interference from hemolysis, lipemia or icterus. "HEM", "LIP", or "ICT" respectively, is printed on the result card in place of the result.
- Extremely elevated amylase levels (>9,000 U/L) will have a significant effect, >10% increase, on the chloride result. The concentration of amylase is not evaluated by the Piccolo system for each specimen.
- The potassium assay in the Piccolo system is a coupled pyruvate kinase (PK) / lactate dehydrogenase (LDH) assay. Therefore, in cases of extreme muscle trauma or highly elevated levels of creatine kinase (CK), the Piccolo may recover a falsely elevated potassium (K<sup>+</sup>) value. In such cases, unexpected high potassium recoveries need to be confirmed utilizing a different methodology.
- For maximum levels of endogenous substances contact Abaxis Technical Support.

### Effects of Exogenous and Therapeutic Substances

- Thirty-five exogenous and therapeutic substances were selected as potential interferents for Abaxis test methods based on recommendations by Young.<sup>15</sup> Significant interference is defined as greater than  $\pm 10\%$  shift in the result for a normal range specimen. Human serum pools were supplemented with known concentrations of the drugs or chemicals and then analyzed. Please see Table 2 for a list of exogenous and therapeutic substances evaluated. **Please see TABLE 3 for a list of analytes where interference was observed.**

**Table 2: Exogenous and Therapeutic Substances Evaluated**

| Potential Interferent             | Highest Concentration Tested<br>(mg/dL unless otherwise specified) |
|-----------------------------------|--|
| Acetaminophen                     | 100  |
| Acetoacetate                      | 102  |
| Acetylsalicylic Acid              | 50   |
| Ampicillin                        | 30   |
| Ascorbic acid                     | 3  |
| Caffeine                          | 10   |
| Cephalothin (Keflin)              | 400  |
| Chloramphenicol                   | 100  |
| Cimetidine                        | 16   |
| Dopamine                          | 13   |
| Epinephrine                       | 1  |
| Erythromycin                      | 10   |
| Glutathione                       | 30   |
| Hydrochlorothiazide               | 7.5  |
| Ibuprofen                         | 50   |
| Isoniazide                        | 4  |
| Ketoprofen                        | 50   |
| L-dopa                            | 5  |
| Lidocaine                         | 1  |
| Lithium Lactate                   | 84   |
| Methicillin                       | 100  |
| Methotrexate                      | 0.5  |
| Metronidazole                     | 5  |
| Nafcillin                         | 1  |
| Nitrofurantoin                    | 20   |
| Oxacillin                         | 1  |
| Oxaloacetate                      | 132  |
| Penicillin G                      | 100  |
| Phenytoin (5,5-Diphenylhydantion) | 3  |
| Proline                           | 4  |
| Rifampin                          | 0.5  |
| Salicylic Acid                    | 50   |
| Sulfadiazine                      | 150  |
| Sulfanilamide                     | 50   |
| Theophylline                      | 20   |

Please see Table 3 for a list of analytes where interference was observed.

**Table 3: The following substances showed greater than  $\pm 10$  % shift in the result for a normal range specimen.**

|                             | Concentration Which<br>Produces > 10%<br>Interference | % Interference <sup>A</sup><br>Observed |
|-----------------------------|---|---|
| <b>Potassium</b>            |   |   |
| Penicillin G                | 100   | 17% inc.                                |
| Sulfadiazine                | 150   | 12% dec.                                |
| <b>Sodium</b>               |   |   |
| Cephalothin                 | 400   | 12% inc.                                |
| Methotrexate                | 0.5   | 11% inc.                                |
| Penicillin G                | 100   | 10% inc.                                |
| <b>Total Carbon Dioxide</b> |   |   |
| Acetaminophen               | 100   | 11% inc.                                |
| Ascorbic Acid               | 20  | 12% dec.                                |
| Cephalothin                 | 400   | 13% inc.                                |
| Cimetidine                  | 16  | 19% dec.                                |
| Erythromycin                | 10  | 21% dec.                                |
| Lidocaine                   | 1   | 23% inc.                                |
| Methotrexate                | 0.5   | 80% dec.                                |
| Nitrofurantoin              | 20  | 13% inc.                                |
| Salicylic Acid              | 50  | 17% dec.                                |
| Sulfadiazine                | 150   | 25% dec.                                |

<sup>A</sup> Dec. = decreased concentration of the specified analyte; Inc. = increased concentration of the specified analyte

- For the Chloride assay, bromide at toxic levels ( $\geq 15$  mmol/L) can cause a significant effect ( $> 10\%$  increase), on the chloride result. Iodide at very high concentrations (30 mmol/L, highest level tested) has no effect. Normal physiological levels of bromide and iodide do not interfere with the Piccolo Chloride Test System.

## 11. Expected Values

Samples from approximately 140 adult males and females were analyzed on the Piccolo blood chemistry analyzer to determine the reference interval. These ranges were calculated based on the 95% reference interval estimated from the combined (overall) values obtained from the reference subjects.<sup>16</sup> It is recommended that your office or institution establish normal ranges for your particular patient population.

**Table 4: Piccolo Reference Intervals**

| Analyte                                       | Common Units   | SI Units       |
|---|----------------|----------------|
| <b>Chloride (CL<sup>-</sup>)</b>              | 98-108 mmol/L  | 98-108 mmol/L  |
| <b>Potassium (K<sup>+</sup>)</b>              | 3.6-5.1 mmol/L | 3.6-5.1 mmol/L |
| <b>Sodium (NA<sup>+</sup>)</b>                | 128-145 mmol/L | 128-145 mmol/L |
| <b>Total Carbon Dioxide (tCO<sub>2</sub>)</b> | 18-33 mmol/L   | 18-33 mmol/L   |

## 12. Performance Characteristics

### Linearity

The chemistry for each analyte is linear over the dynamic range listed below when the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer is operated according to the recommended procedure (refer to the Piccolo blood chemistry analyzer or the Piccolo Xpress chemistry analyzer Operator's Manual).

**Table 5: Piccolo Dynamic Ranges**

| Analyte                                       | Common Units   | SI Units       |
|---|----------------|----------------|
| <b>Chloride (CL<sup>-</sup>)</b>              | 80-135 mmol/L  | 80-135 mmol/L  |
| <b>Potassium (K<sup>+</sup>)</b>              | 1.5-8.5 mmol/L | 1.5-8.5 mmol/L |
| <b>Sodium (NA<sup>+</sup>)</b>                | 110-170 mmol/L | 110-170 mmol/L |
| <b>Total Carbon Dioxide (tCO<sub>2</sub>)</b> | 5-40 mmol/L    | 5-40 mmol/L    |

If the analyte concentration is above the measuring range (dynamic range), but less than the system range, the print card will indicate a ">" sign at the upper limit and an asterisk after the number, e.g. CL<sup>-</sup> >135\* mmol/L. If lower than the dynamic range, a "<" will be printed with an asterisk, e.g. CL<sup>-</sup> <80\* U/L. For values that are grossly beyond the measurement range (system range), "~~~~" will be printed instead of a result. Any time "~~~~" appears on a print card, collect a new sample and rerun the test. If results for the second sample are suppressed again, please call Abaxis Technical Support.

#### Sensitivity (Limits of Detection)

The lower limit of the reportable (dynamic) range for each analyte is: chloride 80 mmol/L; potassium 1.5 mmol/L; sodium 110 mmol/L; and total carbon dioxide 5 mmol/L.

#### Precision

Precision studies were conducted using CLSI EP5-A guidelines<sup>17</sup> with modifications based on CLSI EP18-P<sup>18</sup> for unit-use devices. Results for within-run and total precision were determined using two levels of commercially available control materials. The studies made use of multiple instruments and two reagent disc lots. Potassium and total carbon dioxide testing was performed at two sites over 20 days; sodium testing was performed at one site over 20 days; chloride testing was done at two sites over a period of five days.

Results of precision studies are shown in Table 6.

**Table 6: Precision**

| Analyte                   | Sample Size | Within-Run | Total |
|---------------------------|-------------|------------|-------|
| <b>Chloride (mmol/L)</b>  |             |            |       |
| <u>Control 1</u>          | N = 160     |            |       |
| Mean                      |             | 97.8       | 97.8  |
| SD                        |             | 1.63       | 1.74  |
| CV                        |             | 1.7        | 1.7   |
| <u>Control 2</u>          |             |            |       |
| Mean                      |             | 113.6      | 113.6 |
| SD                        |             | 1.97       | 2.22  |
| CV                        |             | 1.7        | 2.0   |
| <b>Potassium (mmol/L)</b> |             |            |       |
| <u>Control 1</u>          | N = 120     |            |       |
| Mean                      |             | 6.12       | 6.12  |
| SD                        |             | 0.32       | 0.32  |
| CV                        |             | 5.2        | 5.7   |
| <u>Control 2</u>          |             |            |       |
| Mean                      |             | 4.10       | 4.10  |
| SD                        |             | 0.24       | 0.26  |
| CV                        |             | 5.9        | 6.3   |
| <b>Sodium (mmol/L)</b>    |             |            |       |
| <u>Control 1</u>          | N = 80      |            |       |
| Mean                      |             | 143.5      | 143.5 |
| SD                        |             | 2.28       | 2.28  |
| CV                        |             | 1.6        | 1.6   |
| <u>Control 2</u>          |             |            |       |
| Mean                      |             | 120.0      | 120.0 |
| SD                        |             | 2.13       | 2.13  |
| CV                        |             | 1.8        | 1.8   |

**Table 6: Precision (continued)**

| Analyte                              | Sample Size | Within-Run | Total |
|--------------------------------------|-------------|------------|-------|
| <b>Total Carbon Dioxide (mmol/L)</b> |             |            |       |
| <u>Control 1</u>                     | N = 120     |            |       |
| Mean                                 |             | 21.4       | 21.4  |
| SD                                   |             | 2.29       | 2.29  |
| CV                                   |             | 10.7       | 10.7  |
| <u>Control 2</u>                     |             |            |       |
| Mean                                 |             | 10.5       | 10.5  |
| SD                                   |             | 0.90       | 0.90  |
| CV                                   |             | 8.6        | 8.6   |

**Correlation**

Serum samples were collected and assayed on the Piccolo blood chemistry analyzer and by comparative methods. In some cases, high and low supplemented samples were used to cover the dynamic range. The samples were chosen to meet the distribution values in CLSI EP9-A guideline.<sup>19</sup> Representative correlation statistics are shown in Table 7.

**Table 7: Correlation of Piccolo Blood Chemistry Analyzer with Comparative Method(s)**

|                                      | Correlation Coefficient | Slope | Intercept | SEE  | N   | Sample Range (mmol/L) | Comparative Method |
|--------------------------------------|-------------------------|-------|-----------|------|-----|-----------------------|--------------------|
| <b>Chloride (mmol/L)</b>             | 0.978                   | 0.982 | -1.1      | 1.84 | 120 | 71-118                | Vitros 950         |
| <b>Potassium (mmol/L)</b>            | 0.969                   | 0.863 | 0.6       | 0.14 | 58  | 2.0-6.8               | Radiometer KNA™ 2  |
| <b>Sodium (mmol/L)</b>               | 0.937                   | 0.782 | 27.7      | 3.79 | 113 | 116-154               | Radiometer KNA™ 2  |
| <b>Total Carbon Dioxide (mmol/L)</b> | 0.947                   | 0.903 | 2.4       | 0.84 | 60  | 6-39                  | Cobas Fara         |

### Results of Untrained User Study

An “untrained user” study was conducted in which participants were given only the test instructions and asked to perform testing of 3 discs with blinded randomized samples. The samples consisted of serum pools prepared at three levels for each of the four analytes: chloride, potassium, sodium, and total carbon dioxide. The participants were not given any training on the use of the test. A total of approximately 60 participants were enrolled from 3 sites, representing a diverse demographic (educational, age, gender, etc) population.

Tables below present the summary of the performance for each analyte.

#### Chloride (CL<sup>-</sup>)

|  | Level 1                                  | Level 2                                  | Level 3                                  |
|--|--|--|--|
| N  | 62                                       | 62                                       | 62                                       |
| Mean   | 94.6                                     | 106.0                                    | 115.5                                    |
| %CV  | 1.8                                      | 1.4                                      | 1.5                                      |
| Observed Range                               | 90 – 100                                 | 102 – 108                                | 110 – 119                                |
| Percent of Results<br>in the Range<br>± 2.4% | 91.9%<br>57/62<br>95% CI: 82.2% to 97.3% | 96.8%<br>60/62<br>95% CI: 88.8% to 99.6% | 95.2%<br>59/62<br>95% CI: 86.5% to 99.0% |

\* This percent is based on the premise that one cannot distinguish properly between normal and abnormal values when errors are greater than one-quarter of the normal range. The range of (98 mmol/L - 108 mmol/L) was considered.

#### Potassium (K<sup>+</sup>)

|  | Level 1                                | Level 2                                | Level 3                                |
|--|--|--|--|
| N  | 62                                     | 62                                     | 62                                     |
| Mean   | 3.4                                    | 5.7                                    | 7.2                                    |
| %CV  | 3.3                                    | 2.5                                    | 2.0                                    |
| Observed Range                               | 3.2 – 3.7                              | 5.2 – 5.9                              | 6.7 – 7.5                              |
| Percent of Results<br>in the Range<br>± 8.6% | 100%<br>62/62<br>95% CI: 94.2% to 100% | 100%<br>62/62<br>95% CI: 94.2% to 100% | 100%<br>62/62<br>95% CI: 94.2% to 100% |

#### Sodium (NA<sup>+</sup>)

|  | Level 1                                 | Level 2                                | Level 3                                |
|--|---|--|--|
| N  | 62                                      | 62                                     | 62                                     |
| Mean   | 122.1                                   | 140.8                                  | 157.5                                  |
| %CV  | 1.0                                     | 0.8                                    | 1.0                                    |
| Observed Range                               | 118 – 127                               | 138 – 143                              | 154 – 162                              |
| Percent of Results<br>in the Range<br>± 3.1% | 98.4%<br>61/62<br>95% CI: 91.3% to 100% | 100%<br>62/62<br>95% CI: 94.2% to 100% | 100%<br>62/62<br>95% CI: 94.2% to 100% |

#### Total Carbon Dioxide (tCO<sub>2</sub>)

|   | Level 1                                | Level 2                                 | Level 3                                |
|---|--|---|--|
| N   | 62                                     | 62                                      | 62                                     |
| Mean  | 20.3                                   | 27.6                                    | 34.4                                   |
| %CV   | 5.1                                    | 4.6                                     | 3.7                                    |
| Observed Range                                | 18 – 23                                | 23 – 30                                 | 32 – 38                                |
| Percent of Results<br>in the Range<br>± 14.7% | 100%<br>62/62<br>95% CI: 94.2% to 100% | 98.4%<br>61/62<br>95% CI: 91.3% to 100% | 100%<br>62/62<br>95% CI: 94.2% to 100% |

### 13. Bibliography

1. Ono T, et al. A new enzymatic assay of chloride in serum. Clin Chem 1988;34:552-3.
2. Berry MN, et al. Enzymatic determination of potassium in serum. Clin Chem 1989;35:817-20.
3. Van Pelt J. Enzymatic determination of sodium, potassium and chloride in serum compared with determination by flame photometry, coulometry and ion selective electrodes. Clin Chem 1994;40:846-7.
4. Hubl W, et al. Enzymatic determination of sodium, potassium and chloride in abnormal (hemolyzed, icteric, lipemic, paraproteinemic, or uremic) serum samples compared with indirect determination with ion selective electrodes. Clin Chem 1994;40:1528-31.
5. Helgersson RC, et al. Host-guest Complexation. 50. Potassium and sodium ion-selective chromogenic ionophores. J Amer Chem Soc 1989;111:6339-50.
6. Kumar A, et al. Chromogenic ionophore-based methods for spectrophotometric assay of sodium and potassium in serum and plasma. Clin Chem 1988;34:1709-12.
7. Berry MN, et al. Enzymatic determination of sodium in serum. Clin Chem 1988;34:2295-8.
8. Skeggs LT Jr. An automatic method for the determination of carbon dioxide in blood plasma. Am J. Clin Pathol 1960;33:181-5.
9. Korzun WJ, Miller WG. Carbon Dioxide. In: Kaplan LA, Pesce AJ, eds. Clinical chemistry theory, analysis and correlation, 2<sup>nd</sup> ed. St. Louis: The CV Mosby Company, 1989:869-72.
10. CLSI. Physician's office laboratory guidelines, tentative guideline, 2<sup>nd</sup> ed. CLSI Document POL1-T2. Wayne, PA: CLSI, 1992.
11. Scott, M.G. Electrolytes and Blood Gases. In: Burtis CA, Ashwood ER, eds. Tietz Textbook of Clinical Chemistry. 3<sup>rd</sup> ed. Philadelphia: WB Saunders Company, 1999: 1058-9.
12. CLSI. Procedures for the handling and processing of blood specimens; tentative standard. CLSI Document H18-T. Wayne, PA: CLSI, 1984.
13. Scott, M.G. Electrolytes and Blood Gases. In: Burtis CA, Ashwood ER, eds. Tietz Textbook of Clinical Chemistry. 3<sup>rd</sup> ed. Philadelphia: WB Saunders Company, 1999: 1065-6.
14. CLSI. Interference testing in clinical chemistry; proposed guideline. CLSI Document EP7-P. Wayne, PA: CLSI, 1986.
15. Young DS. Effects of drugs on clinical laboratory tests, 3rd ed. Washington, DC: AACC Press, 1990. CLSI.
16. How to define and determine reference intervals in the clinical laboratory, approved guidelines, 2<sup>nd</sup> ed. CLSI Document C28-A2. Wayne, PA: CLSI, 2000.
17. CLSI. Evaluation of precision performance of clinical chemistry devices; approved guideline. CLSI Document EP5-A. Wayne, PA: CLSI, 1999.
18. CLSI. Quality management for unit-use testing; proposed guideline. CLSI Document EP18-P. Wayne, PA: CLSI, 1999.
19. CLSI. Method comparison and bias estimation using patient samples; approved guideline. CLSI Document EP9-A. Wayne, PA: CLSI, 1995.