



Abbott

Release Date: 11/25/24

REF	
GTIN	Product Name

07P6901	<i>Alinity i Free T3 Calibrators</i>
----------------	--------------------------------------

Components:

07P69A	Alinity i Free T3 Calibrator A
07P69B 07P69C 07P69D 07P69E 07P69F	Alinity i Free T3 Calibrators B-F

Abbott Customers:

For additional information, please contact your Abbott Customer Support Center Representative by calling 1-800-527-1869, 1-800-323-9100, or 1-800-235-5396.

Abbott employees:

For additional information relative to the content of the SDSs, please contact your local Safety Representative.

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

1 Identification

- **Product name:** Alinity i Free T3 Calibrator A
- **ADD List number:** 07P69A
- **Application of the substance / mixture:** For In Vitro Diagnostic Use
- **Manufacturer / Supplier:**
Abbott Diagnostics
100 Abbott Park Road
Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT
- **Department issuing SDS:** Abbott Diagnostics Environmental Health and Safety
- **Emergency telephone number**
Contact the CHEMTREC® Emergency Call Center for assistance with transportation or hazardous materials emergencies (24 hours/day, 7 days/week). Refer to Abbott customer number 675805.
Telephone (800) 424-9300 (toll-free) if you are calling from within the United States, Canada, Puerto Rico and the Virgin Islands.

2 Hazard(s) identification

- **Classification of the substance or mixture**
The classification was made according to U.S. OSHA 29 CFR 1910.1200 and 1910.1030 and applicable European regulations, and is expanded upon from supplier company and/or literature data.
This product has been evaluated per the classification criteria in the Globally Harmonized System of Classification and Labelling of Chemicals (GHS). This product does not meet the criteria for classification in accordance with the GHS.
- **Label elements**
- **GHS label elements:** The product is labelled according to the Globally Harmonized System (GHS).
- **Hazard pictograms:** none
- **Signal word:** none
- **Hazard-determining components of labeling:**
Sodium azide
- **Hazard statements:** none
- **Precautionary statements:**
P501 Dispose of contents / container in accordance with local regulations.
- **Routes of Exposure:**
For bloodborne pathogens and potentially infectious materials:
 - non-intact skin
 - mucous membranes (which includes, but is not limited to, the lining of the nose, mouth and throat)
 - parenteral contact (e.g. by injection, puncture)
- **Health:** No adverse effects expected if used as directed.

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

- **Fire:** Noncombustible
- **Reactivity:** Minimal hazard - Stable, even in a fire. Not reactive with water. Not an oxidizer.
- **Other hazards**
This product contains potentially infectious material. Refer to the US OSHA Bloodborne pathogens standard (29 CFR 1910.1030) for additional relevant information.

3 Composition/information on ingredients

- **Chemical characterization:** Mixture of chemical and/or biological substances for in vitro diagnostic use.

Hazardous chemical ingredients per U.S. OSHA criteria (29 CFR 1910.1200 Hazard Communication):

CAS: 26628-22-8	Sodium azide	0.1953%
-----------------	--------------	---------

4 First-aid measures

- **After inhalation:** Remove from source of exposure. Seek medical attention and appropriate follow-up.
- **After skin contact:**
Take off any clothing that the product touched. Wash affected area with soap and water. Seek medical attention and appropriate follow-up.
- **After eye contact:**
Rinse open eye(s) cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Seek medical attention and appropriate follow-up. Wash hands after handling.
- **After swallowing:** Rinse mouth with water. Seek medical attention and appropriate follow-up.
- **Information for Medical Personnel**
This product contains human-sourced and/or potentially infectious material. No known test method can offer complete assurance that products derived from human sources or inactivated microorganisms will not transmit infection. Therefore, all human-sourced material should be considered potentially infectious.

The human-sourced material used in this product has been tested and found to be:
 - Nonreactive for HBsAg (hepatitis B surface antigen)
 - Nonreactive for anti-HCV (antibodies to hepatitis C virus)
 - Nonreactive for HIV-1 Ag or HIV-1 RNA (human immunodeficiency virus type 1 antigen or human immunodeficiency virus type 1 ribonucleic acid)
 - Nonreactive for anti-HIV-1 (antibodies to human immunodeficiency virus type 1)
 - Nonreactive for anti-HIV-2 (antibodies to human immunodeficiency virus type 2)
- **Most important symptoms and effects, both acute and delayed:** None expected
- **Medical conditions aggravated by exposure:** None known

5 Fire-fighting measures

- **Suitable extinguishing agents**
Dry chemical, carbon dioxide (CO₂), water spray or regular foam.
 - Caution: CO₂ will displace air in confined spaces and may cause an oxygen-deficient atmosphere.
 - For larger fires: There are no unique chemical or reactivity hazards that would impact firefighting decisions related

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

to this product. Use firefighting measures that suit the environment.

· **Special hazards arising from the substance or mixture**

There are no unique chemical or reactivity hazards that would impact firefighting decisions due to the chemicals in this product.

No further relevant information available.

· **Protective equipment**

For large fires, wear appropriate heat- and flame-resistant personal protective equipment and a NFPA/NIOSH approved positive-pressure, self-contained breathing apparatus.

6 Accidental release measures

· **Personal precautions, protective equipment and emergency procedures**

Handle as a potentially infectious material.

Minimize exposure by using appropriate personal protective equipment as listed in Section 8. Stop leak if possible.

Keep unprotected persons away.

· **Environmental precautions**

Prevent liquid and vapor from entering sewage system, storm drains, surface waters, and soil.

· **Methods and material for containment and cleaning up**

Blot up small volumes of spilled or spattered product with paper towels or similar materials.

- Contain larger spills by placing absorbents around the outside edges of the spill. Absorb with any material suitable for water-based liquids - e.g. paper towels, universal sorbents, sand, diatomite, sawdust, etc.

Clean the affected area. Suitable cleaners are:

- warm water and detergent or similar cleansing agent

Apply a suitable disinfectant. Select a disinfectant that is effective against bloodborne infectious agents, as well as other microbial agents that you might expect to be prevalent in your population. A disinfectant that is effective against Mycobacterium tuberculosis is generally effective against all known viruses and non-sporeforming bacteria, and is suitable for most clinical laboratory situations.

NOTE: Commercial disinfectants must be used according to manufacturer directions. Disinfectants are typically hazardous chemicals that react with many chemicals, materials and living tissues. Obtain and review the manufacturer's safety information before using the disinfectant.

This product contains sodium azide, which is toxic and reactive. See Sections 10 and 13 for additional information that could affect handling and disposal of contaminated spill materials.

NOTE FOR LARGE-VOLUME SPILL: This product contains sodium azide, which reacts with acid to liberate hydrazoic acid, a very toxic gas. Select a disinfectant with the following properties if disinfection of materials used to absorb a large volume of spilled product is required:

- Do not use any chemical or product with a pH below 6 to disinfect waste that contains sodium azide. Hydrazoic acid, a toxic gas, will be released when the pH is lower than 6.

- Do not use any chemical or product that contains mercury or any other metal to disinfect waste that contains sodium azide. This will create metal azide compounds, which can be highly explosive under pressure or shock (percussion).

- Select a disinfectant that does not bubble, effervesce or otherwise generate aerosols.

- Do not use excess disinfectant.

- Failure to follow manufacturer's directions may lead to unexpected reactions with the waste.

- Do not use a disinfectant if you do not have the proper facility, equipment and other appropriate protective measures available to work with it safely.

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

Dispose of spilled and contaminated material in accordance with Federal, State, and Local regulations. See Section 13 for information that may impact disposal of materials contaminated with this product.

· Reference to other sections

See Section 7 for information on safe handling.
See Section 8 for information on personal protection equipment.
See Section 13 for disposal information.

7 Handling and storage

- **Precautions for safe handling:** Handle as a potentially infectious material.
- **Information about protection against explosions and fires:** The product is not flammable.
- **Requirements to be met by storerooms and receptacles:** Store only in the original container.
- **Information about storage in one common storage facility:** Store in original packaging.
- **Further information about storage conditions:**
Refer to the package insert or product label for additional information on storage conditions for product quality.

8 Exposure controls/personal protection

· Components with Occupational Exposure Limits

CAS: 26628-22-8 Sodium azide (0.1953 %)

REL Ceiling limit value: 0.3** mg/m³, 0.1* ppm
*as HN₃; **as NaN₃; Skin

TLV Ceiling limit value: 0.29** mg/m³, 0.11* ppm
*as HN₃ vapor **as NaN₃, A4

· General protective and hygienic measures:

Always maintain good housekeeping and follow general precautionary measures. Do not eat, drink or store food and beverages in areas where chemicals or specimens are used. Wash hands before breaks, after handling reagents and specimens, and at the end of the workshift.

Observe universal precautions and other appropriate biosafety practices for handling potentially infectious material.

· Breathing equipment:

Normal use and storage of product - respiratory protection is not necessary if room is well ventilated.

Small-volume spills (e.g. small enough to clean up with a paper towel or small sorbent pad) - respiratory protection should not be necessary if room is well ventilated.

Other unusual conditions (e.g. volume spilled too big to clean up with materials in arm's reach) - Use appropriate NIOSH-approved air-purifying respirator if airborne chemical concentrations may exceed the exposure limit (if any) listed above.

Hazardous Materials Emergencies or Firefighting - use NIOSH/NFPA-approved respiratory protection.

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

- **Hand protection:**

Wear impervious gloves if hand contact with the material is anticipated. Dispose of contaminated gloves after use in accordance with applicable laws and good laboratory practices.

- **Material of gloves and breakthrough time of the glove material:**

The glove material must be suitable for use in a microbiological laboratory and have a measured breakthrough time of at least 30 minutes, such as those with a Class 2 protection index per EN374 (or equivalent standard applicable in your region). NOTE: This recommendation applies only to the product stated in this Safety Data Sheet. When dissolving in or mixing with other substances, contact the supplier of approved gloves.

- **Eye protection:**

Wear safety glasses or other protective eyewear. If splash potential exists, wear full face shield or goggles.

- **Body protection:**

Normal use: protect personal clothing from spatters and small spills. Wear a laboratory coat (or other protective clothing required by your institution).

Larger spills (e.g. that can saturate cloth): wear appropriate water-repellant covering over clothing.

9 Physical and chemical properties

- **General Information**

· Color:	Light yellow
· Odor:	Odorless
· Melting point/Melting range:	Not determined
· Boiling point/Boiling range:	Not determined
· Flammability (solid, gaseous)	Not applicable
· Explosion limits	
· Lower:	Not determined
· Upper:	Not determined
· Flash point	Not applicable
· pH-value at 20 °C (68 °F)	8
· Dynamic:	Not determined
· Solubility in / Miscibility with	
· Water:	Not miscible or difficult to mix
· Vapor pressure:	
· Density at 20 °C (68 °F)	1.024 g/cm ³ (8.5453 lbs/gal)
· Form:	Liquid
· Auto igniting	Product is not self-igniting.
· Danger of explosion	Product does not present an explosion hazard.
· Solids content:	0.0 %
· Evaporation rate:	Not determined

10 Stability and reactivity

- **Thermal decomposition / conditions to be avoided**

No decomposition if used and stored according to specifications.

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

· Possibility of hazardous reactions:

This product contains sodium azide. Sodium azide solutions are reported to:

- react with acids to release hydrazoic acid, a very toxic gas. Higher quantities of hydrazoic acid are released as the solution becomes more acidic (i.e., as the pH of the solution gets lower). Low quantities of hydrazoic acid can be released from sodium azide in water.
- react with certain metals (copper, lead, silver, brass) to form explosive metal azide compounds. Violent explosions have been reported during plumbing work on drain systems containing accumulations of azide on copper, lead, brass, or solder.

· **Conditions to avoid:** No further relevant information available.

· **Hazardous decomposition products:** No dangerous decomposition products known.

11 Toxicological information

- **Acute toxicity** Based on available data, the classification criteria are not met.
- **LD50/LC50 values for hazardous ingredients per OSHA criteria:**
- **Ingredients (100% pure substance/s):** Not applicable.
- **Skin irritation:** Based on available data, the classification criteria are not met.
- **Eye irritation:** Based on available data, the classification criteria are not met.
- **Sensitization:** Based on available data, the classification criteria are not met.
- **Germ cell mutagenicity** Based on available data, the classification criteria are not met.
- **Carcinogenicity** Based on available data, the classification criteria are not met.
- **Reproductive toxicity** Based on available data, the classification criteria are not met.
- **Specific target organ toxicity - single exposure** Based on available data, the classification criteria are not met.
- **Specific target organ toxicity - repeated exposure** Based on available data, the classification criteria are not met.
- **Aspiration hazard** Based on available data, the classification criteria are not met.

· Carcinogenic categories

· IARC (International Agency for Research on Cancer)

None of the ingredients is listed.

· NTP (National Toxicology Program)

None of the ingredients is listed.

· OSHA-Ca (Occupational Safety & Health Administration)

None of the ingredients is listed.

· **Target organs/systems:** Unknown

12 Ecological information

· **Aquatic toxicity:** No further relevant information available.

· Results of PBT and vPvB assessment

- **PBT:** Not applicable
- **vPvB:** Not applicable



Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

- **Additional ecological information**
- **General notes:**
Do not allow undiluted product or large quantities of it to reach ground water, water course, or sewage system.

13 Disposal considerations

- **Recommendation for disposal of unused product:**
Dispose in accordance with federal, state and local regulations and institutional requirements. Waste containing this product may be considered hazardous per U.S. EPA, state or local regulations. The following may be particularly important when identifying appropriate disposal:
 - Potentially infectious. See Section 4, Information for Medical Personnel, for more information.
 - See Section 6 for information when institutional or regulatory requirements include any sort of treatment of potentially infectious waste.
 - Contains sodium azide. See Section 10 when considering how to appropriately dispose of unused product. For drain systems with pipes or solder containing copper, lead, brass and/or silver, flush drains thoroughly with copious amounts of water to prevent the formation of potentially explosive metal azides in plumbing. Detailed information about azides in drains is available from the U.S. NIOSH Current Intelligence Bulletin No. 13 (August 16, 1976).
- **Recommendation for disposal of packaging:**
Non-contaminated packaging may be used for recycling. Refer to applicable local regulations and institutional policies. For disposal of contaminated packaging, refer to applicable local regulations and institutional policies.
- **Recommended cleansing agent:** Water with cleansing agents, if necessary.

14 Transport information

· DOT, ADN, IMDG, IATA none

· **UN proper shipping name**
 · DOT, ADR, ADN, IMDG, IATA none

· **Transport hazard class(es)**
 · DOT, ADR, ADN, IMDG, IATA
 · Class none
 · DOT, IMDG, IATA none

· **Environmental hazards**
 · Marine pollutant: No

· **Additional information**

· DOT
 · Remarks: Not restricted for transportation.
 · IMDG
 · Remarks: Not restricted for transportation.



Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

- IATA
- **Remarks:** Not restricted for transportation.

15 Regulatory information

· **SARA (Superfund Amendments and Reauthorization Act of 1986 - USA):**

· **Section 302/304 (40CFR355.30 / 40CFR355.40):**

CAS: 26628-22-8 Sodium azide

· **Section 313 (40CFR372.65):**

CAS: 26628-22-8 Sodium azide

· **Hazardous Air Pollutants**

None of the ingredients is listed.

· **California Proposition 65 (USA):**

· **Chemicals known to cause cancer:**

The product does not contain listed substances.

· **Chemicals known to cause female reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause male reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause developmental reproductive toxicity:**

None of the ingredients is listed.

16 Other information

The information and recommendations contained herein are based upon information or tests believed to be reliable. Abbott Laboratories does not guarantee the accuracy or completeness of this information or recommendations contained herein, NOR SHALL ANY OF THIS INFORMATION CONSTITUTE A WARRANTY, WHETHER EXPRESSED OR IMPLIED, AS TO THE SAFETY OF THE GOODS, THE MERCHANTABILITY OF THE GOODS, OR THE FITNESS OF THE GOODS FOR A PARTICULAR PURPOSE.

This information is not a substitute for the advice of a health care professional, nor is it a recommendation for any particular course of treatment. It is not intended to supplement, modify or supersede any information (e.g. labeling and package inserts) provided with respect to the medical use of the product. Abbott Laboratories assumes no responsibility for results obtained or for incidental or consequential damages, including lost profits, arising from the use of these data. No warranty against infringement of any patent, copyright or trademark is made or implied.

The information in this Safety Data Sheet (SDS) reflects the most current hazard information for this product.

· **Department issuing SDS**

- Abbott Diagnostics Safety, Health and Environmental Assurance
Department 0571



Abbott

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrator A

· **Contact**

- General information about this product:

Abbott Diagnostics
Technical Support
100 Abbott Park Road
Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT

· **Date of preparation / last revision 11/25/2024 / 44**

· **Abbreviations and acronyms:**

ADR: Accord européen sur le transport des marchandises dangereuses par Route (European Agreement concerning the International Carriage of Dangerous Goods by Road)
IMDG: International Maritime Code for Dangerous Goods
DOT: US Department of Transportation
IATA: International Air Transport Association
EINECS: European Inventory of Existing Commercial Chemical Substances
ELINCS: European List of Notified Chemical Substances
CAS: Chemical Abstracts Service (Division of the American Chemical Society)
LC50: Lethal concentration, 50 percent
LD50: Lethal dose, 50 percent
PBT: persistent, bioaccumulative and toxic
vPvB: very persistent and very bioaccumulative
OSHA: Occupational Safety & Health
TLV: Threshold Limit Value
PEL: Permissible Exposure Limit
REL: Recommended Exposure Limit

· *** Sections marked with an asterisk (*) have been altered since the previous version.**

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

1 Identification

- **Product name:** Alinity i Free T3 Calibrators B-F
- **ADD List number:**
 - 07P69B
 - 07P69C
 - 07P69D
 - 07P69E
 - 07P69F
- **Application of the substance / mixture:** For In Vitro Diagnostic Use
- **Manufacturer / Supplier:**
 - Abbott Diagnostics
 - 100 Abbott Park Road
 - Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT
- **Department issuing SDS:** Abbott Diagnostics Environmental Health and Safety
- **Emergency telephone number**

Contact the CHEMTREC® Emergency Call Center for assistance with transportation or hazardous materials emergencies (24 hours/day, 7 days/week). Refer to Abbott customer number 675805.
Telephone (800) 424-9300 (toll-free) if you are calling from within the United States, Canada, Puerto Rico and the Virgin Islands.

2 Hazard(s) identification

- **Classification of the substance or mixture**

The classification was made according to U.S. OSHA 29 CFR 1910.1200 and 1910.1030 and applicable European regulations, and is expanded upon from supplier company and/or literature data.

This product has been evaluated per the classification criteria in the Globally Harmonized System of Classification and Labelling of Chemicals (GHS). This product does not meet the criteria for classification in accordance with the GHS.
- **Label elements**
- **GHS label elements:** The product is labelled according to the Globally Harmonized System (GHS).
- **Hazard pictograms:** none
- **Signal word:** none
- **Hazard-determining components of labeling:**
 - Sodium azide
- **Hazard statements:** none
- **Precautionary statements:**
 - P501 Dispose of contents / container in accordance with local regulations.

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

· Routes of Exposure:

For bloodborne pathogens and potentially infectious materials:

- non-intact skin
- mucous membranes (which includes, but is not limited to, the lining of the nose, mouth and throat)
- parenteral contact (e.g. by injection, puncture)

· **Health:** No adverse effects expected if used as directed.

· **Fire:** Noncombustible

· **Reactivity:** Minimal hazard - Stable, even in a fire. Not reactive with water. Not an oxidizer.

· Other hazards

This product contains potentially infectious material. Refer to the US OSHA Bloodborne pathogens standard (29 CFR 1910.1030) for additional relevant information.

3 Composition/information on ingredients

· **Chemical characterization:** Mixture of chemical and/or biological substances for in vitro diagnostic use.

· **Hazardous chemical ingredients per U.S. OSHA criteria (29 CFR 1910.1200 Hazard Communication):**

CAS: 26628-22-8	Sodium azide	0.1897%
-----------------	--------------	---------

4 First-aid measures

· **After inhalation:** Remove from source of exposure. Seek medical attention and appropriate follow-up.

· **After skin contact:**

Take off any clothing that the product touched. Wash affected area with soap and water. Seek medical attention and appropriate follow-up.

· **After eye contact:**

Rinse open eye(s) cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Seek medical attention and appropriate follow-up. Wash hands after handling.

· **After swallowing:** Rinse mouth with water. Seek medical attention and appropriate follow-up.

· **Information for Medical Personnel**

This product contains human-sourced and/or potentially infectious material. No known test method can offer complete assurance that products derived from human sources or inactivated microorganisms will not transmit infection. Therefore, all human-sourced material should be considered potentially infectious.

The human-sourced material used in this product has been tested and found to be:

- Nonreactive for HBsAg (hepatitis B surface antigen)
- Nonreactive for anti-HCV (antibodies to hepatitis C virus)
- Nonreactive for HIV-1 Ag or HIV-1 RNA (human immunodeficiency virus type 1 antigen or human immunodeficiency virus type 1 ribonucleic acid)
- Nonreactive for anti-HIV-1 (antibodies to human immunodeficiency virus type 1)
- Nonreactive for anti-HIV-2 (antibodies to human immunodeficiency virus type 2)

· **Most important symptoms and effects, both acute and delayed:** None expected

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

· **Medical conditions aggravated by exposure:** None known

5 Fire-fighting measures

· **Suitable extinguishing agents**

Dry chemical, carbon dioxide (CO₂), water spray or regular foam.

- Caution: CO₂ will displace air in confined spaces and may cause an oxygen-deficient atmosphere.
- For larger fires: There are no unique chemical or reactivity hazards that would impact firefighting decisions related to this product. Use firefighting measures that suit the environment.

· **Special hazards arising from the substance or mixture**

There are no unique chemical or reactivity hazards that would impact firefighting decisions due to the chemicals in this product.

No further relevant information available.

· **Protective equipment**

For large fires, wear appropriate heat- and flame-resistant personal protective equipment and a NFPA/NIOSH approved positive-pressure, self-contained breathing apparatus.

6 Accidental release measures

· **Personal precautions, protective equipment and emergency procedures**

Handle as a potentially infectious material.

Minimize exposure by using appropriate personal protective equipment as listed in Section 8. Stop leak if possible. Keep unprotected persons away.

· **Environmental precautions**

Prevent liquid and vapor from entering sewage system, storm drains, surface waters, and soil.

· **Methods and material for containment and cleaning up**

Blot up small volumes of spilled or spattered product with paper towels or similar materials.

- Contain larger spills by placing absorbents around the outside edges of the spill. Absorb with any material suitable for water-based liquids - e.g. paper towels, universal sorbents, sand, diatomite, sawdust, etc.

Clean the affected area. Suitable cleaners are:

- warm water and detergent or similar cleansing agent

Apply a suitable disinfectant. Select a disinfectant that is effective against bloodborne infectious agents, as well as other microbial agents that you might expect to be prevalent in your population. A disinfectant that is effective against Mycobacterium tuberculosis is generally effective against all known viruses and non-sporeforming bacteria, and is suitable for most clinical laboratory situations.

NOTE: Commercial disinfectants must be used according to manufacturer directions. Disinfectants are typically hazardous chemicals that react with many chemicals, materials and living tissues. Obtain and review the manufacturer's safety information before using the disinfectant.

This product contains sodium azide, which is toxic and reactive. See Sections 10 and 13 for additional information that could affect handling and disposal of contaminated spill materials.

NOTE FOR LARGE-VOLUME SPILL: This product contains sodium azide, which reacts with acid to liberate hydrazoic acid, a very toxic gas. Select a disinfectant with the following properties if disinfection of materials used to absorb a large volume of spilled product is required:

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

- Do not use any chemical or product with a pH below 6 to disinfect waste that contains sodium azide. Hydrazoic acid, a toxic gas, will be released when the pH is lower than 6.
- Do not use any chemical or product that contains mercury or any other metal to disinfect waste that contains sodium azide. This will create metal azide compounds, which can be highly explosive under pressure or shock (percussion).
- Select a disinfectant that does not bubble, effervesce or otherwise generate aerosols.
- Do not use excess disinfectant.
- Failure to follow manufacturer's directions may lead to unexpected reactions with the waste.
- Do not use a disinfectant if you do not have the proper facility, equipment and other appropriate protective measures available to work with it safely.

Dispose of spilled and contaminated material in accordance with Federal, State, and Local regulations. See Section 13 for information that may impact disposal of materials contaminated with this product.

Reference to other sections

- See Section 7 for information on safe handling.
- See Section 8 for information on personal protection equipment.
- See Section 13 for disposal information.

7 Handling and storage

- **Precautions for safe handling:** Handle as a potentially infectious material.
- **Information about protection against explosions and fires:** The product is not flammable.
- **Requirements to be met by storerooms and receptacles:** Store only in the original container.
- **Information about storage in one common storage facility:** Store in original packaging.
- **Further information about storage conditions:**
Refer to the package insert or product label for additional information on storage conditions for product quality.

8 Exposure controls/personal protection

Components with Occupational Exposure Limits

CAS: 26628-22-8 Sodium azide (0.1897 %)

REL Ceiling limit value: 0.3** mg/m³, 0.1* ppm
*as HN₃; **as NaN₃; Skin

TLV Ceiling limit value: 0.29** mg/m³, 0.11* ppm
*as HN₃ vapor **as NaN₃, A4

General protective and hygienic measures:

Always maintain good housekeeping and follow general precautionary measures. Do not eat, drink or store food and beverages in areas where chemicals or specimens are used. Wash hands before breaks, after handling reagents and specimens, and at the end of the workshift.

Observe universal precautions and other appropriate biosafety practices for handling potentially infectious material.

Breathing equipment:

Normal use and storage of product - respiratory protection is not necessary if room is well ventilated.

Small-volume spills (e.g. small enough to clean up with a paper towel or small sorbent pad) - respiratory protection

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

should not be necessary if room is well ventilated.

Other unusual conditions (e.g. volume spilled too big to clean up with materials in arm's reach) - Use appropriate NIOSH-approved air-purifying respirator if airborne chemical concentrations may exceed the exposure limit (if any) listed above.

Hazardous Materials Emergencies or Firefighting - use NIOSH/NFPA-approved respiratory protection.

· **Hand protection:**

Wear impervious gloves if hand contact with the material is anticipated. Dispose of contaminated gloves after use in accordance with applicable laws and good laboratory practices.

· **Material of gloves and breakthrough time of the glove material:**

The glove material must be suitable for use in a microbiological laboratory and have a measured breakthrough time of at least 30 minutes, such as those with a Class 2 protection index per EN374 (or equivalent standard applicable in your region). NOTE: This recommendation applies only to the product stated in this Safety Data Sheet. When dissolving in or mixing with other substances, contact the supplier of approved gloves.

· **Eye protection:**

Wear safety glasses or other protective eyewear. If splash potential exists, wear full face shield or goggles.

· **Body protection:**

Normal use: protect personal clothing from spatters and small spills. Wear a laboratory coat (or other protective clothing required by your institution).

Larger spills (e.g. that can saturate cloth): wear appropriate water-repellant covering over clothing.

9 Physical and chemical properties

· **General Information**

· Color:	Light yellow
· Odor:	Odorless
· Melting point/Melting range:	Not determined
· Boiling point/Boiling range:	Not determined
· Flammability (solid, gaseous)	Not applicable
· Explosion limits	
· Lower:	Not determined
· Upper:	Not determined
· Flash point	Not applicable
· pH-value at 20 °C (68 °F)	8
· Dynamic:	Not determined
· Solubility in / Miscibility with	
· Water:	Not miscible or difficult to mix
· Vapor pressure:	
· Density at 20 °C (68 °F)	1.027 g/cm ³ (8.5703 lbs/gal)
· Form:	Liquid
· Auto igniting	Product is not self-igniting.
· Danger of explosion	Product does not present an explosion hazard.
· Water:	3.8 %



Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

- **Solids content:** 0.0 %
- **Evaporation rate:** Not determined

10 Stability and reactivity

- **Thermal decomposition / conditions to be avoided**
No decomposition if used and stored according to specifications.
- **Possibility of hazardous reactions:**
This product contains sodium azide. Sodium azide solutions are reported to:
 - react with acids to release hydrazoic acid, a very toxic gas. Higher quantities of hydrazoic acid are released as the solution becomes more acidic (i.e., as the pH of the solution gets lower). Low quantities of hydrazoic acid can be released from sodium azide in water.
 - react with certain metals (copper, lead, silver, brass) to form explosive metal azide compounds. Violent explosions have been reported during plumbing work on drain systems containing accumulations of azide on copper, lead, brass, or solder.
- **Conditions to avoid:** No further relevant information available.
- **Hazardous decomposition products:** No dangerous decomposition products known.

11 Toxicological information

- **Acute toxicity** Based on available data, the classification criteria are not met.
- **LD50/LC50 values for hazardous ingredients per OSHA criteria:**
- **Ingredients (100% pure substance/s):** Not applicable.
- **Skin irritation:** Based on available data, the classification criteria are not met.
- **Eye irritation:** Based on available data, the classification criteria are not met.
- **Sensitization:** Based on available data, the classification criteria are not met.
- **Germ cell mutagenicity** Based on available data, the classification criteria are not met.
- **Carcinogenicity** Based on available data, the classification criteria are not met.
- **Reproductive toxicity** Based on available data, the classification criteria are not met.
- **Specific target organ toxicity - single exposure** Based on available data, the classification criteria are not met.
- **Specific target organ toxicity - repeated exposure** Based on available data, the classification criteria are not met.
- **Aspiration hazard** Based on available data, the classification criteria are not met.

· Carcinogenic categories

· IARC (International Agency for Research on Cancer)

None of the ingredients is listed.

· NTP (National Toxicology Program)

None of the ingredients is listed.

· OSHA-Ca (Occupational Safety & Health Administration)

None of the ingredients is listed.

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

· **Target organs/systems:** Unknown

12 Ecological information

- **Aquatic toxicity:** No further relevant information available.
- **Results of PBT and vPvB assessment**
 - **PBT:** Not applicable
 - **vPvB:** Not applicable
- **Additional ecological information**
- **General notes:**
Do not allow undiluted product or large quantities of it to reach ground water, water course, or sewage system.

13 Disposal considerations

- **Recommendation for disposal of unused product:**
Dispose in accordance with federal, state and local regulations and institutional requirements. Waste containing this product may be considered hazardous per U.S. EPA, state or local regulations. The following may be particularly important when identifying appropriate disposal:
 - Potentially infectious. See Section 4, Information for Medical Personnel, for more information.
 - See Section 6 for information when institutional or regulatory requirements include any sort of treatment of potentially infectious waste.
 - Contains sodium azide. See Section 10 when considering how to appropriately dispose of unused product. For drain systems with pipes or solder containing copper, lead, brass and/or silver, flush drains thoroughly with copious amounts of water to prevent the formation of potentially explosive metal azides in plumbing. Detailed information about azides in drains is available from the U.S. NIOSH Current Intelligence Bulletin No. 13 (August 16, 1976).
- **Recommendation for disposal of packaging:**
Non-contaminated packaging may be used for recycling. Refer to applicable local regulations and institutional policies. For disposal of contaminated packaging, refer to applicable local regulations and institutional policies.
- **Recommended cleansing agent:** Water with cleansing agents, if necessary.

14 Transport information

- **DOT, ADN, IMDG, IATA** none
- **UN proper shipping name**
 - **DOT, ADR, ADN, IMDG, IATA** none
- **Transport hazard class(es)**
 - **DOT, ADR, ADN, IMDG, IATA**
 - **Class** none
 - **DOT, IMDG, IATA** none



Abbott

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

· Environmental hazards

· Marine pollutant: No

· Additional information

· DOT

· Remarks: Not restricted for transportation.

· IMDG

· Remarks: Not restricted for transportation.

· IATA

· Remarks: Not restricted for transportation.

15 Regulatory information

· SARA (Superfund Amendments and Reauthorization Act of 1986 - USA):

· Section 302/304 (40CFR355.30 / 40CFR355.40):

CAS: 26628-22-8 Sodium azide

· Section 313 (40CFR372.65):

CAS: 26628-22-8 Sodium azide

· Hazardous Air Pollutants

None of the ingredients is listed.

· California Proposition 65 (USA):

· Chemicals known to cause cancer:

The product does not contain listed substances.

· Chemicals known to cause female reproductive toxicity:

None of the ingredients is listed.

· Chemicals known to cause male reproductive toxicity:

None of the ingredients is listed.

· Chemicals known to cause developmental reproductive toxicity:

None of the ingredients is listed.

16 Other information

The information and recommendations contained herein are based upon information or tests believed to be reliable. Abbott Laboratories does not guarantee the accuracy or completeness of this information or recommendations contained herein, NOR SHALL ANY OF THIS INFORMATION CONSTITUTE A WARRANTY, WHETHER EXPRESSED OR IMPLIED, AS TO THE SAFETY OF THE GOODS, THE MERCHANTABILITY OF THE GOODS, OR THE FITNESS OF THE GOODS FOR A PARTICULAR PURPOSE.

This information is not a substitute for the advice of a health care professional, nor is it a recommendation for any particular course of treatment. It is not intended to supplement, modify or supersede any information (e.g. labeling and package inserts) provided with respect to the medical use of the product. Abbott Laboratories assumes no responsibility for results obtained or for incidental or consequential damages, including lost profits, arising from the use

Safety Data Sheet

according to OSHA Hazard Communication standard 29CFR 1910.1200

© Abbott Laboratories Release date 11/25/2024

Last alteration on 11/25/2024

Product name: Alinity i Free T3 Calibrators B-F

of these data. No warranty against infringement of any patent, copyright or trademark is made or implied.

The information in this Safety Data Sheet (SDS) reflects the most current hazard information for this product.

· **Department issuing SDS**

- Abbott Diagnostics Safety, Health and Environmental Assurance
Department 0571

· **Contact**

- General information about this product:
Abbott Diagnostics
Technical Support
100 Abbott Park Road
Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT

· **Date of preparation / last revision 11/25/2024 / 43**

· **Abbreviations and acronyms:**

ADR: Accord européen sur le transport des marchandises dangereuses par Route (European Agreement concerning the International Carriage of Dangerous Goods by Road)

IMDG: International Maritime Code for Dangerous Goods

DOT: US Department of Transportation

IATA: International Air Transport Association

EINECS: European Inventory of Existing Commercial Chemical Substances

ELINCS: European List of Notified Chemical Substances

CAS: Chemical Abstracts Service (Division of the American Chemical Society)

LC50: Lethal concentration, 50 percent

LD50: Lethal dose, 50 percent

PBT: persistent, bioaccumulative and toxic

vPvB: very persistent and very bioaccumulative

OSHA: Occupational Safety & Health

TLV: Threshold Limit Value

PEL: Permissible Exposure Limit

REL: Recommended Exposure Limit

· *** Sections marked with an asterisk (*) have been altered since the previous version.**