

Micro-Touch® NitraFree™

Micro-Touch® NitraFree™ pink nitrile examination gloves are manufactured without sulfur-based chemical accelerators,* helping to protect you from Type I (latex) and Type IV (chemical) allergies. Its premium nitrile material provides excellent protection in all risk settings, and has more stretch for greater comfort and fit. Proceeds from the purchase of every case go to Susan G. Komen for the Cure® in support of breast cancer research and outreach.†

Safety & Protection

Infection Control

Allergy Concerns. This examination glove is made of 100% nitrile. It does not contain sulfur-based chemical accelerators or natural rubber latex, thus preventing Type I allergies and reducing the potential for Type IV allergic reactions.


Sterilization. This is a non-sterile examination glove.

Quality Benchmarks. Meets or exceeds ASTM examination glove standards. Manufactured within the quality guidelines of ISO 13485:2003, ISO 9001:2000, and FDA-QSR. FDA approved for handling chemotherapy drugs.¹

Cuff. Inverted, 9.5" beaded cuff.

Resistance. Nitrile film has an excellent puncture and chemical resistance. Nitrile gloves are recommended for conditions where high strength and chemical protection are required.

This glove is tested for use with the following 9 chemotherapy drugs: Vincristine Sulfate, Cyclophosphamide, Doxorubicin Hydrochloride, 5-Fluorouracil, Methotrexate, Etoposide, Paclitaxel, ThioTEPA, and Carmustine.²



Skin Protection

Powder-Free. To help eliminate powder-related complications and sensitivity.

Comfort & Fit

Hand Fatigue

Stretchability (Modulus). Nitrile has high elasticity, providing wearers with good comfort and fit.

Shape of Glove/Former. Ambidextrous.

Tactile Sensitivity

Dependability. The softer, stretch formulation of this glove provides superior tactile sensitivity in a synthetic glove.

Memory. Nitrile has high elasticity and good memory, allowing the film to adapt to the wearer's hand.

Grip

Application Driven. The textured fingertips of the glove provide extra protection for wet or dry grip, and against slippage.

Ordering Information

SIZE	Product Number
X-Small	6034510
Small	6034511
Medium	6034512
Large	6034513
X-Large	6034514



100 Gloves/Dispenser Box – 10 Boxes/Case

* Does not contain any sulfur-based accelerators such as thiurams, carbamates, thiazoles, guanidine or thioureas.

† In 2010, Ansell will donate \$2 per case sold of the Micro-Touch® NitraFree™ to Susan G. Komen for the Cure®, with a guaranteed minimum donation of \$25,000.

1. Specific chemical resistance data available upon request: 1-800-952-9916 (U.S.A. only). Gloves used for protection against chemotherapy drugs must be selected specifically for the type of chemicals used. Review material safety data sheets for the chemicals being used to determine the required level of protection.

2. CAUTION: Testing showed an average breakthrough time of 45.3 minutes with Carmustine.

MICRO-TOUCH®
NITRAFREE™



Proud Partner
Susan G. Komen for the Cure®

Ansell

Everything you touch...we touch.

Material 100% nitrile rubber (acrylonitrile / butadiene / methacrylic acid terpolymer); no natural rubber or sulfur-based chemical accelerators*

Dimensional Specifications

SIZE	Minimum Length mm (inches)	Average Palm Width mm (inches)
X-Small	240 (9.4)	75 (3.0)
Small	240 (9.4)	85 (3.3)
Medium	240 (9.4)	95 (3.8)
Large	240 (9.4)	106 (4.2)
X-Large	240 (9.4)	116 (4.6)

Thickness (single)

Data from Sampling	Average Thickness
Fingertip, mm (mils)	0.08 (3.1)
Palm, mm (mils)	0.06 (2.4)

Physical Properties (ASTM, EN, and ISO Requirements)

	Before Aging	After Aging
Freedom from Holes (inspection level 1)	2.5 AQL**	N/A
Tensile, MPa (p.s.i.) minimum	14.0 (2,030)	14.0 (2,030)
Elongation (%) minimum	500	400

Ansell manufacturing specifications meet or exceed ASTM specifications.

Color	Pink
Accelerators	None
Certification	Manufacturing facility registered to ISO 9001 and ISO 13485 FDA approved for handling chemotherapy drugs ¹
Donning Agent	None (powder-free)
Residual Powder Content	<2.0 mg/glove
Quality Control	In compliance with ASTM D6319 examination glove standards
References	ANSI/ASQ Z1.4 Sampling Procedures and Tables for Inspection by Attributes
Country of Origin	Malaysia
Storage	Cool environment, shielded from ozone and UV light

Except for applicable ASTM requirements, the above information represents typical data for the product and should not be construed or used as an actual product performance specification.

*Does not contain any sulfur-based accelerators such as thiurams, carbamates, thiazoles, guanidine or thioureas

**Acceptable quality level – maximum % of gloves with defects

1. Specific chemical resistance data available upon request: 1-800-952-9916 (U.S.A. only). Gloves used for protection against chemotherapy drugs must be selected specifically for the type of chemicals used. Review material safety data sheets for the chemicals being used to determine the required level of protection.

For additional information and samples, call Ansell Healthcare at 800-952-9916 or visit us at www.ansellhealthcare.com

Ansell Healthcare Products LLC | ansellhealthcare.com
200 Schulz Drive | Red Bank, NJ 07701 U.S.A. | Phone: (800) 952-9916 (U.S.A. only)

Ansell, Micro-Touch, and NitraFree are trademarks owned by Ansell Limited or one of its affiliates. ©2010 All Rights Reserved.
The COLOR PINK® and PINK NITRILE® are registered trademarks owned by Color World, LLC. 2006 All Rights Reserved.

SS460 02/2010

Printed on recycled paper using soy inks.



Ansell

Everything you touch...we touch.