

Application Instructions for Metal Buckle:



1. With the Posey label on the outside of the belt and the belt around the patient's waist, pass the metal tip of the belt through the buckle and over the teeth of the buckle.



2. Bring tip of belt across front of buckle and slip it through the other side. Adjust the belt so it is snug, but not uncomfortable for the patient. You should be able to slip your open hand (flat) between the belt and the patient.



3. Strap should lay flat across buckle. Tuck excess under belt. Always verify proper closure before use.



Always grasp the transfer belt from underneath. Assist patient using your facility's approved procedures.

Application Instructions for the Posey Quick-Release Buckle:

1. Release the quick-release buckle by pushing inward on the quick-release buttons.
2. Wrap the buckle around the patient's wrist and snap the buckle closed. Pull the strap to tighten. Adjust the belt so it is snug, but not uncomfortable for the patient. Make sure you can slide your open hand (flat) between the belt and the patient.
3. Turn the buckle around to the patient's back (Tuck excess strap under belt).



⚠ WARNING The quick release buckle should be positioned so that the patient can not release the buckle during transfer. Release of the buckle during transfer may result in injury. **NEVER leave patient unattended with product applied!**

NOTE: THIS PRODUCT IS NOT INTENDED TO BE USED AS A RESTRAINT!

The Posey Company recommends the user of this product check and follow its facility's written procedures for approved assisted walking, lifting, and/or transfer procedures.