



SportNeb™ 2

Compressor Nebulizer

3050-2



Table of Contents

IEC Symbols.....	3
Important Safeguards.....	4
Introduction.....	4
Specifications.....	5
Description.....	6
Operation Instructions.....	7
Cleaning.....	7
Maintenance.....	7
Warranty.....	8

IEC Symbols



Caution, Attention or Consult accompanying documents



Alternating Current



Type BF Equipment



Double Insulation



Single Use Equipment

Important Safeguards



Caution:

Federal Law (USA) restricts this device for sale or use on the order of a physician. Follow the instructions or you physician to operate this unit.



Danger: To reduce the risk of electrocution:

1. Always unplug the unit immediately after using.
2. Do not use while bathing.
3. Do not place or store the unit where it can fall or be pulled into a tub or sink.
4. Do not place or drop into water or other liquid.
5. Do not reach for a product that has fallen into water. Unplug immediately.



Warning: To reduce the risk of burns, electrocution, fire or injury:

1. Electrical Shock Hazard - do not remove the cabinet or open the cover.
2. Disconnect the power cord from the electrical outlet before cleaning or servicing.
3. Do not place this equipment near hot, sparking or burning objects.

Important Safeguards




Warning: To reduce the risk of burns, electrocution, fire or injury:

4. Do not use oil or grease on or near this device.
5. Turn the unit off when it is not in use.
6. Keep the cord away from **HEATED** or **HOT** surfaces.
7. **NEVER** drop or insert any object into any opening.
8. **NEVER** block the air openings of the product or place it on a soft surface, such as a bed or couch, where the air opening may be blocked.
9. Avoid the operation in wet or damp locations.
10. Unplug the unit before filling the nebulizer.



Warning: To reduce the risk of infection:

1. The Nebulizer kit with a  mark on it should not be reused.
2. To prevent the risk of cross infection, different persons should never use the same nebulizer kit.
3. If you are using a reusable nebulizer kit, cleaning of the nebulizer is recommended after each aerosol treatment. Disinfecting is recommended once a day. Since nebulizers vary in type and construction, we recommend that you follow the cleaning instructions provided with the nebulizer. If there is no cleaning instruction accompanying with the nebulizer, follow the recommended cleaning procedures described on page 6.
4. Use only the FDA or CE approved nebulizer kits.

Introduction

This compressor nebulizer, 3050-2, is designed to efficiently deliver the prescribed saline solution to treat your respiratory disorders, such as asthma, allergies, and bronchitis. The SportNeb2 will convert the saline solution into an aerosol mist that penetrates through your lungs.


Read this manual thoroughly before using nebulizer and save for future reference.

Specification

Electrical Requirements	120VAC/60Hz
Power Consumption	110W Max
Dimensions	4.1" (105mm) High ×5.7" (145mm) Wide × 7.9" (200mm) Deep
Weight	1.6kg (3.53lbs)
Sound Pressure Level	58dB(A)
Max. Comp. Pressure	30 Psig or greater
Neb. Operating Pressure	10 Psig or greater
Compressor Free Air Flow	8lpm or greater
Particle Size	Less Than 5 microns
Operating Temperature	10°C ~ 40°C (50°F to 104°F)
Operating Humidity	Up to 95% non-condensing
Storage/Transport Temperature Range	-20°C ~ 60°C (-4°F to 140°F)
Storage/Transport Humidity	Up to 95% non-condensing

- (1). Conditions may vary based on altitude above sea level, barometric pressure and temperature. Specification subjects to change without notice.
- (2). This product complies with the following standards: C22.2 No 601.1-M90, UL Std. No 2601-1 and IEC/EN 60601-1-1.
- (3). For electromagnetic compatibility the standards are based as follow: CSA 601.1-2 and IEC/EN 60601-1-2.
- (4). Protection against electric shock is classified as **Type II**.
- (5). Degree of protection against electric shock is classified as **Type BF**.
- (6). Mode of operation belongs to continuous mode.
- (7). Equipment not suitable for use in the presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide
- (8). Degree of protection against ingress of water is rated as IPX0.

Description



- 1. Air Outlet Connector
- 2. Filter Cap
- 3. Power Switch
- 4. Power Cord
- 5. Nebulizer T-Piece

- 6. Mouthpiece
- 7. Nebulizer Cap
- 8. Baffle
- 9. Nebulizer Chamber
- 10. Air Inlet Connector

- 11. Tubing

Operation Instructions

1. Place compressor on a stable, sturdy and flat surface such that the unit can be easily reached when you're seated.
2. Pull out the power cord from the base of the unit, and make sure that the unit is in the "off" (0) position by pressing on the left side of switch.
3. Plug power cord into a 120VAC/60Hz wall outlet.
4. Connect one end of the nebulizer tubing into the air outlet connector.
5. Add the prescribed saline solution through the opening on cap using an eye dropper or per-measured dose container.
6. Assemble mouthpiece and T-piece, and insert into the top of the nebulizer cap. If an aerosol mask applied connect the bottom of mask to the top of nebulizer cap.
7. Connect tubing to nebulizer air inlet connector.
8. Turn the power on to start the unit.
9. Implement the treatment by placing the mouthpiece between the teeth. Inhale and exhale through the mouthpiece.
10. If aerosol applied place the mask over the mouth and nose.

Cleaning

1. Turn the power off and unplug from the wall outlet.
2. Remove the tubing from the air inlet.
3. Disassemble mouthpiece, T-piece, nebulizer cap, chamber and baffle, and wash these items except the tubing in hot water with dish-washing detergent. Rinse these items thoroughly to remove the detergent.
4. Mix one part of white vinegar with 3 part of hot distilled water, and use a clean container to carry this disinfectant solution. Submerging mouthpiece, T-piece, nebulizer cap, chamber and baffle for half an hour
5. Remove these cleaned units from container and let them dry in the air.
6. There is no need to clean the tubing if necessary wipe the surface of it regularly.

Maintenance

1. The filter of air compressor should be replaced when it turns gray. The additional filter can be procured from your dealer
2. Authorized personnel maintain the rest of this machine only.

Warranty

Drive Medical warrants the SportNeb2 nebulizer compressor to be free from defective workmanship and materials for a period of 5 years from the date of purchase. Check with your dealer for more details. This warranty is limited to the original purchaser of the compressor. Any defective part(s) will be replaced, at the sole discretion and determination of Drive Medical, if the unit has not been misused or tampered with during the warranty period. Normal maintenance items, as outlined in this manual, and disposable components are not covered by this warranty. Shipping charges, if any, shall be paid by the purchaser.

NOTE:

There is no other express warranty. Implied warranties, including those of merchantability and fitness for a particular purpose, are limited to the duration of the express limited warranty and to the extent permitted by law and all implied warranties are excluded.



Drive Medical

99 Seaview Boulevard
Port Washington, NY 11050
Toll Free: 877.224.0946
Local: 516.998.4600
www.drivemedical.com